THIRUVANANTHAPURAM REGIONAL CO-OPERATIVE MILK PRODUCERS UNION LTD KOLLAM DAIRY, THEVALLY, KOLLAM - 691 009. GST:32AAAAT9795J4ZU[PAN:AAAAT9795J]

Email:

Enquiry No: 1138/24

Dtd 12-JUL-24.

KGST No. TIN:32020212025 dt CST No.0202C100193 dt 17/03/1986

To

SUPPLIERS/MANUFACURERS/DEALERS (LAB EQUIPMENTS)

Sir.

Subject : QUOTATION FOR HYGIENE MONITORING SYSTEM

Reference:

Thiruvananthapuram Regional Co-Operative Milk Producers Union Ltd, Kollam Dairy, Thevally, Kollam - 691 009., invites your lowest quotations and shortest delivery period for the items as per Annexure. The offer shall be subjected to the following terms and conditions:

Terms & Conditions:

- 1. PRICE BASIS: All prices must be quoted as FOR site basis, including all taxes and duties, freight, packing, forwarding transit risk, loading, etc. for the safe delivery of the items and nothing extra will be paid on this account.
- 2. VALIDITY: The offer should be valid for 2 MONTHS for acceptance from the date of your offer.
- 3. GUARANTEE: The items supplied to be guaranteed for performance and quality for one year. If any item found damaged, bad quality etc. to be replaced free of cost.
- 4. PAYMENT: 100% payment will be released on satisfactory receipt of items at our site after satisfying the sample results.
- 5. LIQUIDATED DAMAGES CLAUSE: In the case of supply is not completed within the stipulated period, liquidated damage @1% will be charged for the items so delayed subject to a maximum of 10% of the total value.
- 6. The undersigned reserves the right to accept your offer partly/or in full or to reject without giving any explanation.
- 7. The offer should be sent in a sealed cover superscribed : ENQUIRY FOR HYGIENE MONITORING SYSTEM due on 25/07/2024

addressed to:

SENIOR MANAGER

THIRUVANANTHAPURAM REGIONAL CO-OPERATIVE MILK PRODUCERS UNION LTD THEVALLY, KOLLAM - 691 009.

Yours faithfully,

Encl: Annexure

Emal 15-7-24

THIRUVANANTHAPURAM REGIONAL CO-OPERATIVE MILK PRODUCERS UNION LTD KOLLAM DAIRY, THEVALLY, KOLLAM - 691 009. GST:32AAAAT9795J4ZU[PAN:AAAAT9795J]

Make

Email:

Enquiry No: 1138/24

Dtd 12-JUL-24.

Annexure

The System/Instrument should be simple to use, flexible and accurate quality monitoring system along with features to collect, analyse and report data from multiple quality indicators

Technical Specification

Sl No Item code & Description

Operating Temp Range 5 degree C to 40 degree C

Relative Humidity Range 20% to 85% non-condensing

Measurement Range 0 to 9999 RLUs

Measurement Resolution 1 RLU

Measurement Time 15 seconds approx

Measurement noise +/-5% or +/- RLUs

Memory min for 2000 tests

Battery Backup Required

Serial Interface EIA - 232 campatible

AOAC Aprroval Required

End of Report

Manager Relam Dairy

Quantity Required Unit

1.00 Nos